

STERILIZATION OF THE DS1921

When using the 1921 in food analysis, we need to sterilize the button. The normal process is done at 121°C during 15 minutes.

We've done it once, and the button still run correctly. (we do not measure temperature during sterilization)

Can we proceed that way as a regular basis?

If not, what would be the maximal temperature + time allowed as a regular basis?

Has someone been faced with this question?

rgds

Benoist MACE
Sales Manager
PROGES PLUS
Tel : +33 3 20 64 63 63
Fax : +33 3 20 64 63 64
b.mace.at.proges.com
www.proges.com

Benoist Mace b.mace.at.proges.com

Mon, 9 Oct 2000 16:24:40